

CERTIFICATE

Survey of 12 September 2025

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Virus Genome Detection - Human Herpesvirus 6 (HHV-6) (405):

Validity 12 months:

HHV-6 (DNA) - quantitative

HHV-6 (DNA) - qualitative

(R) analysis is subject to the RiliBÄK

Customer:

10073534


Centrum Immunologii
Centrum Onkologii Ziemi Lubelskiej
im. św. Jana z Dukli
dr K. Jaczewskiego 7
20-090 Lublin



Düsseldorf, 06 October 2025



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. med. Jörg Timm
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 12 September 2025

You have participated in the External Quality Assessment with the following analysis

Virus Genome Detection - Human Herpesvirus 6 (HHV-6) (405):

HHV-6 (DNA) - quantitative

HHV-6 (DNA) - qualitative

(R) analysis is subject to the RiliBÄK

Customer:

10073534


**Centrum Immunologii
Centrum Onkologii Ziemi Lubelskiej
im. św. Jana z Dukli
dr K. Jaczewskiego 7
20-090 Lublin**



Düsseldorf, 06 October 2025



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. med. Jörg Timm
(Adviser)

Listing and evaluation of the results

10073534: Centrum Immunologii
Centrum Onkologii Ziemi Lubelskiej
im. św. Jana z Dukli

Survey of 12 September 2025

Adviser: Prof. Dr. med. Jörg Timm
Universitätsklinikum Düsseldorf
Heinrich-Heine-Universität
Moorenstr. 5
40225 Düsseldorf


Tel.: +49 211-81-12225
Mail: joerg.timm@med.uni-duesseldorf.de

405

Virus Genome Detection - Human Herpesvirus 6 (HHV-6)

HHV-6 (DNA) - quantitative

Name of test	Unit	Method	Manufacturer				
HHV-6/7 PCR Kit	Copies/ml	PCR / NAT (395)	GENEPROOF (GP)				
Result (quant)	Sample 405A- 250826	Your value	Lower limit	Upper limit	TV- Type	Target value	Meets criteria
	-01	6194	1837	183710	SV	18371	+
	-02	0.000	0.000	0.000	SV	0.000	+
	-03	2848	728	72790	SV	7279	+
	-04	17905	5793	579250	SV	57925	+



HHV-6 (DNA) - qualitative

Name of test	Unit	Method	Manufacturer				
HHV-6/7 PCR Kit		PCR / NAT (395)	GENEPROOF (GP)				
Result (qual)	Sample 405A- 250826	Your specification(s)	Correct specification(s)		TV- Type		Meets criteria
	-01	positive (3)	positive (3)		SV		+
	-02	below detection limit/negative (1)	below detection limit/negative (1)		SV		+
	-03	positive (3)	positive (3)		SV		+
	-04	positive (3)	positive (3)		SV		+

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".